

Equipment Covers

Endo Camera

CODE	DESCRIPTION (WxL)	CASE QTY	CASE DIMENSIONS
N 44-0596CC	Camera Cover/Laser Drape, Closed System Coupler, 5x96" (13x244cm)	100	20x16x12" (52x40x31cm)
S 44-0596CC-S	Camera Cover/Laser Drape, Closed System Coupler, 5x96" (13x244cm)	20	20x16x12" (52x40x31cm)
N 44-0796CC	Camera Cover/Laser Drape, Closed System Coupler, 7x96" (18x244cm)	100	20x16x12" (52x40x31cm)
S 44-0796CC-S	Camera Cover/Laser Drape, Closed System Coupler, 7x96" (18x244cm)	20	20x16x12" (52x40x31cm)
N 44-0996CC	Camera Cover/Laser Drape, Closed System Coupler, 9x96" (23x244cm)	100	20x16x12" (52x40x31cm)
S 44-0996CC-S	Camera Cover/Laser Drape, Closed System Coupler, 9x96" (23x244cm)	20	20x16x12" (52x40x31cm)

Materials

- Clear Polyethylene (PE) Film
- Ethylene Methyl Acrylate Copolymer
- Acrylonitrile Butadiene Styrene (ABS)
- Nylon Agraffe

Product Attributes

- Closed system coupler
- Sterile application stick
- Conformable plastic tip

Sterilization

Sterile: Sterilized by EO (ethylene oxide). Product is sterile unless the package has been damaged or opened.

Non-Sterile: If a product is delivered non-sterile and is ultimately for use in sterile environments, it is intended to be sterilized by the re-packager. This includes further packaging and sterilization according to the validated processes of the re-packager. These products are to be sterilized by ethylene oxide (EO). Other methods of sterilization have not been validated by Alleset.

Shelf Life

Sterile: 5 years

Non-Sterile: Established at 5 years based upon sterile packaged product aging studies and normal warehouse storage conditions. After 5 years, the product should be performance tested to ensure compliance to specifications.

Product Certifications/Standards

- CE certified as Class I, according to MDR 2017/745 Annex VIII, Rule 1

Manufacturing Certifications/Registrations

- EN ISO 13485:2016
- Medical Device Single Audit Process Certified

Storage

- General warehouse conditions



Alleset
4142 Industry Way
Flowery Branch, GA 30542,
USA
www.alleset.com
SM-TDS-003 2024
Country of Origin Abbreviations per ISO 3166-1, Alpha-2 Codes.



GRI Medical & Electronic Technology Co., LTD
1805 Honggao Road,
314031 Jiaxing, Zhejiang
PEOPLE'S REPUBLIC OF CHINA
www.alleset.com
US FDA Establishment Registration #3004911384



Emergo Europe B.V.
Westervoortsedijk 60,
6827 AT Arnhem,
The Netherlands.
www.emergobyul.com

