CODE

Equipment Covers Endo Camera

DESCRIPTION (WxL)



CASE DIMENSIONS

CASE QTY

	OODL			OAGE QTT		
	E44-06120SE	Camera Cover, Sewn Elastic Tip, 5x119" (13x302c	em)	100	15x12x16" (39x30x40cm)	
S	E44-06120SE-S	Camera Cover, Sewn Elastic Tip, 5x119" (13x302c	em)	40	19x15x16" (47x38x30cm)	
Mat	terials					
○ Clear Polyethylene (PE) Film				5in (13cm)		
Product Attributes				ſ		
	wn elastic tip rdboard sleeve op	ening				
Ste	rilization					
Sterile: Sterilized by EO (ethylene oxide). Product is sterile unless the package has been damaged or opened. Non-Sterile: If a product is delivered non-sterile and is ultimately for use in sterile environments, it is intended to be sterilized by the re- packager. This includes further packaging and sterilization according to the validated processes of the re-packager. These products are to be sterilized by ethylene oxide (EO). Other methods of sterilization have not been validated by Alleset.				119in (30	2cm)	
She	elf Life					
Non proc Afte	duct aging studies	ed at 5 years based upon sterile packaged and normal warehouse storage conditions. luct should be performance tested to ensure cations.				
Pro	duct Certificatio	ons/Standards				
	erile: CE certified a Ile 1	as Class I, according to MDD93/42/EEC Annex IX,		ļ	white	
				<u> </u>		

Manufacturing Certifications/Registrations

Non-Sterile: CE certified as Class I, according to MDR 2017/745

• EN ISO 13485:2016

Annex VIII, Rule 1

Medical Device Single Audit Process Certified

Storage

· General warehouse conditions



Alleset 4142 Industry Way Flowery Branch, GA 30542, USA www.alleset.com SM-TDS-003 2023 Country of Origin Abbreviations per ISO 3166-1, Alpha-2 Codes.



GRI Medical & Electronic Technology Co., LTD 1805 Honggao Road, 314031 Jiaxing, Zhejiang PEOPLE'S REPUBLIC OF CHINA www.alleset.com US FDA Establishment Registration #3004911384



Emergo Europe B.V. Westervoortsedijk 60, 6827 AT Arnhem, The Netherlands www.alleset.com

