

# Equipment Covers

## ECO<sup>2</sup> Endo Camera

CODE	DESCRIPTION (WxL)	CASE QTY	CASE DIMENSIONS
E44-0596P	ECO <sup>2</sup> Camera Cover, Perforated Tip, 5x96" (13x244cm)	100	13x11x7" (34x27x19cm)
E44-0596P-S	ECO <sup>2</sup> Camera Cover, Perforated Tip, 5x96" (13x244cm)	50	13x11x7" (34x27x19cm)
E44-0796P	ECO <sup>2</sup> Camera Cover, Perforated Tip, 7x96" (18x244cm)	100	11x9x10" (28x23x25cm)
E44-0796P-S	ECO <sup>2</sup> Camera Cover, Perforated Tip, 7x96" (18x244cm)	50	11x9x10" (28x23x25cm)

### Materials

- Clear Polyethylene (PE) Film

### Product Attributes

- Perforated film tip
- Adhesive polyethylene securement strap
- Incorporates sustainable materials which enhance product biodegradation.

### Sterilization

**Sterile:** Sterilized by EO (ethylene oxide). Product is sterile unless the package has been damaged or opened.  
**Non-Sterile:** If a product is delivered non-sterile and is ultimately for use in sterile environments, it is intended to be sterilized by the re-packager. This includes further packaging and sterilization according to the validated processes of the re-packager. These products are to be sterilized by ethylene oxide (EO). Other methods of sterilization have not been validated by Alleset.

### Shelf Life

**Sterile:** 5 years  
**Non-Sterile:** Established at 5 years based upon sterile packaged product aging studies and normal warehouse storage conditions. After 5 years, the product should be performance tested to ensure compliance to specifications.

### Product Certifications/Standards


- Sterile: CE certified as Class I, according to MDD93/42/EEC Annex IX, Rule 1
- Non-Sterile: CE certified as Class I, according to MDR 2017/745 Annex VIII, Rule 1

### Manufacturing Certifications/Registrations

- EN ISO 13485:2016
- Medical Device Single Audit Process Certified

### Storage

- General warehouse conditions



Alleset  
 4142 Industry Way  
 Flowery Branch, GA 30542,  
 USA  
[www.alleset.com](http://www.alleset.com)  
 SM-TDS-003 2023  
 Country of Origin Abbreviations per ISO 3166-1, Alpha-2 Codes.

GRI Medical & Electronic Technology Co., LTD  
 1805 Honggao Road,  
 314031 Jiaxing, Zhejiang  
 PEOPLE'S REPUBLIC OF CHINA  
[www.alleset.com](http://www.alleset.com)  
 US FDA Establishment Registration #3004911384

Emergo Europe B.V.  
 Westervoortsedijk 60,  
 6827 AT Arnhem,  
 The Netherlands  
[www.alleset.com](http://www.alleset.com)

