

Surgical Organization

Diathermy Instrument Pouches

CODE	DESCRIPTION	CASE QTY	CASE DIMENSIONS
S 80-14804-S	Diathermy Instrument Pouch, 2 Compartment, 12x16" (32x40cm)	100	12x7x11" (32x17x27cm)
N 80-14805	Diathermy Instrument Pouch, 2 Compartment, 13x16" (33x40cm)	500	24x17x6" (60x44x15cm)
S 80-14805-S	Diathermy Instrument Pouch, 2 Compartment, 13x16" (33x40cm)	100	12x9x12" (31x23x30cm)
N 80-14806	Diathermy Instrument Pouch, 3 Compartment, 12x20" (30x50cm)	500	24x9x12" (60x23x30cm)
S 80-14806-S	Diathermy Instrument Pouch, 3 Compartment, 12x20" (30x50cm)	100	12x9x12" (31x23x30cm)
N 80-14807	Diathermy Instrument Pouch, 1 Compartment, 13x11" (33x28cm)	500	24x9x12" (60x23x30cm)
S 80-14807-S	Diathermy Instrument Pouch, 1 Compartment, 13x11" (33x28cm)	100	12x9x12" (31x23x30cm)
N 80-14808	Diathermy Instrument Pouch, 2 Compartment, 14x24" (35x61cm)	200	14x14x11" (36x36x27cm)
N 80-14809	Diathermy Instrument Pouch, 2 Compartment, 12x15" (30x38cm)	TBD	TBD
N 80-14810	Diathermy Instrument Pouch, 2 Compartment, 30x12" (77x31cm)	200	24x9x12" (60x22x30cm)
N 80-14815	Diathermy Instrument Pouch, 1 Compartment, 20x24" (50x60cm)	250	16x12x13" (40x30x33cm)

Materials

- Clear Polyethylene (PE) Film
- Positioning Tape

Product Attributes

- Adhesive edge for secure placement

Sterilization

Sterile: Sterilized by EO (ethylene oxide). Product is sterile unless the package has been damaged or opened
Non-Sterile: If a product is delivered non-sterile and is ultimately for use in sterile environments, it is intended to be sterilized by the re-packer. This includes further packaging and sterilization according to the validated processes of the re-packer. These products are to be sterilized by ethylene oxide (EO). Other methods of sterilization have not been validated by Alleset.

Shelf Life

Sterile: 5 years
Non-Sterile: Established at 5 years based upon sterile packaged product aging studies and normal warehouse storage conditions. After 5 years, the product should be performance tested to ensure compliance to specifications.

Product Certifications/Standards


- Sterile: CE certified as Class I, according to MDD93/42/EEC Annex IX, Rule 1
- Non-Sterile: CE certified as Class I, according to MDR 2017/745 Annex VIII, Rule 1

Manufacturing Certifications/Registrations

- EN ISO 13485:2016

Storage


- General warehouse conditions



Alleset
 4142 Industry Way
 Flowery Branch, GA 30542,
 USA
www.alleset.com
 SM-TDS-003 2023
 Country of Origin Abbreviations per ISO 3166-1, Alpha-2 Codes.



GRI Medical & Electronic Technology Co., LTD
 1805 Honggao Road,
 314031 Jiaxing, Zhejiang
 PEOPLE'S REPUBLIC OF CHINA
www.alleset.com
 US FDA Establishment Registration #3004911384



Emergo Europe B.V.
 Westervoortsedijk 60,
 6827 AT Arnhem,
 The Netherlands
www.emergobyul.com

