

CORE Room Turnaround Kits®

ECO² Drawcord Bag

| CODE | DESCRIPTION (LxW) | CASE QTY | CASE DIMENSIONS |
|------------|--|----------|------------------------|
| E50-DCRD01 | ECO ² Drawcord Bag, Red, 29x38" (74x97cm) | 100 | 16x13x10" (40x32x25cm) |
| E50-DCRD02 | ECO ² Drawcord Bag, Red, 33x38" (84x97cm) | 100 | 16x13x10" (40x32x25cm) |
| E50-DCRD03 | ECO ² Drawcord Bag, Red, 38x38" (97x97cm) | 100 | 16x13x10" (40x32x25cm) |
| E50-DCRD04 | ECO ² Drawcord Bag, Red, 40x40" (102x102cm) | 100 | 16x13x10" (40x32x25cm) |
| E50-DCRD05 | ECO ² Drawcord Bag, Red, 40x34" (102x112cm) | 100 | 22x12x8" (56x31x21cm) |

Materials

- ECO² Red Polyethylene (PE) Film
- ECO² Black Polypropylene (PP) Cord

Product Attributes


- Drawcord closure with cord lock
- Printed bio-hazard warning
- Certain components of this product incorporate ECO² sustainable materials which enhance product biodegradability

Sterilization

Non-Sterile: If a product is delivered non-sterile and is ultimately for use in sterile environments, it is intended to be sterilized by the re-packer. This includes further packaging and sterilization according to the validated processes of the re-packer. These products are to be sterilized by ethylene oxide (EO). Other methods of sterilization have not been validated by Alleset.

Shelf Life

Non-Sterile: Established at 1 year based upon sterile packaged product aging studies and normal warehouse storage conditions. After 1 year, the product should be performance tested to ensure compliance to specifications.



Product Certifications/Standards

- CE certified as Class I, according to MDR 2017/745 Annex VIII, Rule 1
- ASTM D5511-12 Standard Test Method for Determining Anaerobic Biodegradation of Plastics Materials Under High-Solids Anaerobic Digestion Conditions

Manufacturing Certifications/Registrations

- EN ISO 13485:2016

Storage

- General warehouse conditions



Alleset
4142 Industry Way
Flowery Branch, GA 30542,
USA
www.alleset.com
SM-TDS-002 2022
Country of Origin Abbreviations per ISO 3166-1, Alpha-2 Codes.



GRI Medical & Electronic Technology Co., LTD
1805 Honggao Road,
314031 Jiaxing, Zhejiang
PEOPLE'S REPUBLIC OF CHINA
www.alleset.com
US FDA Establishment Registration #3004911384



GRI-Alleset Limited B.V.
Kingsfordweg 151,
1043 GR Amsterdam,
The Netherlands
www.alleset.com

