Surgical Drapes

Mayo Stand Cover





Tri-Laminate Nonwoven Material (PP/PP/PE)

Blue Polyethylene (PE) Film

Product Attributes

· Reinforced area for instrument placement.

Sterilization

Non-Sterile: If a product is delivered non-sterile and is ultimately for use in sterile environments, it is intended to be sterilized by the repackager. This includes further packaging and sterilization according to the validated processes of the re-packager. These products are to be sterilized by ethylene oxide (EO). Other methods of sterilization have not been validated by Alleset.

Shelf Life

Non-Sterile: Established at 5 years based upon sterile packaged product aging studies and normal warehouse storage conditions. After 5 years, the product should be performance tested to ensure compliance to specifications.

Product Certifications/Standards


• CE certified as Class I, according to MDR 2017/745 Annex VIII, Rule 1

Manufacturing Certifications/Registrations

- EN ISO 13485:2016
- Medical Device Single Audit Process Certified

Storage

· General warehouse conditions

















GRI Medical & Electronic Technology Co., LTD 1805 Honggao Road, 314031 Jiaxing, Zhejiang PEOPLE'S REPUBLIC OF CHINA www.alleset.com US FDA Establishment Registration #3004911384 EC REP Eme

Emergo Europe B.V. Westervoortsedijk 60, 6827 AT Arnhem, The Netherlands www.emergobyul.com

