# Surgical Drapes EENT/ Neurology



	CODE	DESCRIPTION	CASE QTY	CASE DIMENSIONS
	80-09144	EENT Split Drape, 76x127" (193x323cm)	30	24x16x18" (60x40x45cm)
S	80-09144-S	EENT Split Drape, 76x127" (193x323cm)	26	24x16x18" (60x40x45cm)

### Materials

- Multilayered SMS Material with Antistatic Properties and Alcohol Repellency
- Hydra<sup>™</sup> Absorbent Reinforcement
- Non-Slip Material

#### **Product Attributes**

- Split: 3x29" (8x74cm)
- Absorbent reinforcement material for fluid management
- Tube holders
- Instrument pad
  Adhesive tape for secure placement

## Sterilization

Sterile: Sterilized by EO (ethylene oxide). Product is sterile unless the package has been damaged or opened.

Non-Sterile: If a product is delivered non-sterile and is ultimately for use in sterile environments, it is intended to be sterilized by the repackager. This includes further packaging and sterilization according to the validated processes of the re-packager. These products are to be sterilized by ethylene oxide (EO). Other methods of sterilization have not been validated by Alleset.

#### Shelf Life

#### Sterile: 5 years

Non-Sterile: Established at 5 years based upon sterile packaged product aging studies and normal warehouse storage conditions. After 5 years, the product should be performance tested to ensure compliance to specifications.

#### **Product Certifications/Standards**

CE certified as Class I, according to MDR 2017/745 Annex VIII, Rule 1

### **Manufacturing Certifications/Registrations**

- EN ISO 13485:2016
- Medical Device Single Audit Process Certified

#### Storage


· General warehouse conditions



Alleset 4142 Industry Way Flowery Branch, GA 30542, USA www.alleset.com SM-TDS-003 2023 Country of Origin Abbreviations per ISO 3166-1, Alpha-2 Codes. THINK.MAKE.PROTECT.



GRI Medical & Electronic Technology Co., LTD 1805 Honggao Road, 314031 Jiaxing, Zhejiang PEOPLE'S REPUBLIC OF CHINA www.alleset.com US FDA Establishment Registration #3004911384





EC REP

