

Surgical Drapes

Basic

CODE	DESCRIPTION (AxB)	CASE QTY	CASE DIMENSIONS
N 80-12401	Half Drape, 44x60" (112x152cm)	88	21x17x11" (54x44x29cm)
S 80-12401-S	Half Drape, 44x60" (112x152cm)	20	14x11x10" (36x29x25cm)
N 80-12401-HB	Half Drape, Handi-Bin, 44x60" (112x152cm)	960	47x31x30" (119x79x77cm)
N 80-12402	Medium Drape, 44x76" (112x193cm)	58	18x11x19" (45x29x48cm)
S 80-12402-S	Medium Drape, 44x76" (112x193cm)	28	18x12x14" (45x30x36cm)
N 80-12403	Large Drape, 60x76" (152x193cm)	48	21x17x11" (54x44x29cm)
S 80-12403-S	Large Drape, 60x76" (152x193cm)	20	18x12x14" (45x30x36cm)
N 80-12403-HB	Large Drape, Handi-Bin, 60x76" (152x193cm)	525	47x31x30" (119x79x77cm)
N 80-12404	X-Large Drape, 76x100" (193x254cm)	36	21x17x11" (54x44x29cm)
S 80-12404-S	X-Large Drape, 76x100" (193x254cm)	30	22x13x20" (56x32x50cm)
N 80-12404-HB	X-Large Drape, Handi-Bin, 76x100" (193x254cm)	375	47x31x30" (119x79x77cm)

Materials

- Multilayered SMS Material with Antistatic Properties and Alcohol Repellency

Product Attributes

- Versatile utility drapes meet a wide variety of procedural needs.

Sterilization

Sterile: Sterilized by EO (ethylene oxide). Product is sterile unless the package has been damaged or opened.
Non-Sterile: If a product is delivered non-sterile and is ultimately for use in sterile environments, it is intended to be sterilized by the re-packer. This includes further packaging and sterilization according to the validated processes of the re-packer. These products are to be sterilized by ethylene oxide (EO). Other methods of sterilization have not been validated by Alleset.

Shelf Life

Sterile: 5 years
Non-Sterile: Established at 5 years based upon sterile packaged product aging studies and normal warehouse storage conditions. After 5 years, the product should be performance tested to ensure compliance to specifications.

Product Certifications/Standards


- CE certified as Class I, according to MDR 2017/745 Annex VIII, Rule 1

Manufacturing Certifications/Registrations

- EN ISO 13485:2016
- Medical Device Single Audit Process Certified

Storage

- General warehouse conditions



Alleset
 4142 Industry Way
 Flowery Branch, GA 30542,
 USA
www.alleset.com
 SM-TDS-003 2023
 Country of Origin Abbreviations per ISO 3166-1, Alpha-2 Codes.



GRI Medical & Electronic Technology Co., LTD
 1805 Honggao Road,
 314031 Jiaxing, Zhejiang
 PEOPLE'S REPUBLIC OF CHINA
www.alleset.com
 US FDA Establishment Registration #3004911384



Emergo Europe B.V.
 Westervoortsedijk 60,
 6827 AT Arnhem,
 The Netherlands
www.emergobyul.com

